



Certificate of Analysis

Product Artificial Urine for Corrosion Testing - Stabilized
Part Number 1700-0016
Primary Application Artificial Urine for corrosion testing of urological implants
Volume 200 mL
Lot Number 2203002
Production Date 02 March 2022
Expiration Date 02 March 2023
Storage Conditions Dry at 8° to 25°C
Acceptance Criteria

1. Appearance:.....PASS
2. pH = 6.0 ± 0.1 @ $25 \pm 0.4^\circ\text{C}$:.....PASS
3. Other components gravimetrically-certified:.....PASS

Components

DI Water	Sodium Phosphate	Sodium Sulfate	Preservative
Sodium Chloride	Calcium Chloride	Sodium Citrate	
Potassium Chloride	Ammonium Chloride	Magnesium Sulfate	

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By: Maria Ofitserova, Ph.D.

Date: 08 March 2022

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED