



## CERTIFICATE OF ANALYSIS

### Artificial Urine for Corrosion Testing, Stabilized 1700-0016

**Product:** Artificial Urine for corrosion testing of urological implants

**Part Number:** 1700-0016

**Lot Number:** 904005

**Production Date:** 02 Apr 2019

**Expiration Date:** 02 Apr 2021

**Primary Application:** Product Testing

**Acceptance Criteria:**

1. pH =  $6.0 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N) y

**Storage Conditions:** Dry at 8 to 25° C

**Analysis Certified by:** Jim Murphy **Date:** April 3, 2019  
President  
Title

Rev. A, 03/26/14