



CERTIFICATE OF ANALYSIS

Artificial Urine DIN EN 1616:1999 1700-0017

Product: Artificial Urine prepared according to DIN EN 1616:1999 standard procedure

Part Number: 1700-0017

Lot Number: 609027

Production Date: Sep 15 2016

Expiration Date: Sep 15 2017

Primary Application: For testing of urethral catheters.

Acceptance Criteria:

1. pH = 6.6 ± 0.1 @ 25 ± 0.4 °C (Y/N)

y

Storage Conditions: Keep refrigerated

Analysis Certified by: Jim Murphy

President

Title

Date:

Sept. 19, 2016

Rev. A, 03/26/14