



CERTIFICATE OF ANALYSIS

Artificial Urine DIN EN 1616:1999, Stabilized 1700-0558

Product: Artificial Urine DIN EN 1616:1999, stabilized

Part Number: 1700-0558

Lot Number: 005008

Production Date: May 5, 2020

Expiration Date: May 5, 2021

Primary Application: For testing of urethral catheters.

Acceptance Criteria:

1. pH = 6.6 ± 0.1 @ 25 ± 0.4 °C (Y/N)

Y

Storage Conditions: Dry at 8 ° to 25° C

Analysis Certified by: Rebecca Smith

Vice President

Title

Date: July 20, 2020

Rev. A, 05/14/19