



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-column Carbamate Pesticides Analysis

CB910

Product: OPA Diluent for Post-column analysis of Carbamate pesticides

Part Number: CB910

Lot Number: 709032

Production Date: 27 Sep 2017

Expiration Date: 27 Sep 2019

Primary Application: Post-column analysis of Carbamate pesticides.

Acceptance Criteria

- | | |
|---|----------|
| 1. pH = 9.10 ± 0.1 @ 25 ± 0.4 °C (Y/N) | <u>y</u> |
| 2. Baseline noise (0 ± 25 mvolts) (Y/N) | <u>y</u> |
| 3. Twelve appropriate peaks in chromatogram (Y/N) | <u>y</u> |

Sample: Carbamate Test Mixture, **PN:** 1700-0063

Storage Conditions: Dry at 8 °C to 25 °C

Analysis Certified by: Jim Murphy **Date:** 29 Sep 2017
President
Title