



CERTIFICATE OF ANALYSIS

Sodium Diluent for Amino Acid Analysis

Na220

Product: Sodium Diluent for Amino Acid Analysis

Part Number: Na220

Lot Number: 007037

Production Date: July 22, 2020

Expiration Date: July 22, 2022

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH = 2.20 ± 0.01 @ 25 ± 0.4 °C (Y/N)

Y

2. Baseline: no peaks (Y/N)

Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Rebecca Smith

Vice President

Title

Date: August 13, 2020