



CERTIFICATE OF ANALYSIS

Column Regenerant for Lithium-based Amino Acids Analysis

RG003

Product: Lithium Column Regenerant

Part Number: RG003

Lot Number: 810023

Production Date: 17 Oct 2018

Expiration Date: 17 Oct 2020

Primary Application: For column regeneration in lithium-based amino acids analysis

Acceptance Criteria:

1. pH = 12.60 ± 0.20 @ 25 ± 0.4 °C (Y/N) y

2. Conductivity meets specifications (Y/N) y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy **Date:** October 19, 2018
President
Title