



CERTIFICATE OF ANALYSIS

TRIONE Ninhydrin Reagent T100/T100C

Product: TRIONE, Ninhydrin Reagent

Part Number: T100/T100C

Lot Number: 2404055

Production Date: 30 April 2024

Expiration Date: 30 August 2024

Primary Application: Amino acid analysis

Acceptance Criteria:

1. TRIONE color without shades of brown or purple (Y/N)
2. L-aspartic acid to L-proline ratio ≥ 3.0 (Y/N)

Y

Y

Storage Conditions: Dry at 8 to 25 °C

Sample: Aspartic acid and Proline mixture.

Analysis Certified by: David Mazawa

Title: QC Manager

Date: 06 May 2024